



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product	91288 - Lanthanum Nitrate 0.1M - [10277-43-7]
Batch No	1015667
Analysis Date	19-Jan-2023
Date of Manufacture	January 2023
Expiry Date/Re-Test Date	January 2026
Molecular Formula	La(NO ₃) ₃ .6H ₂ O
Molecular Weight	433.01

Test Parameters	Standards	Actual Results
Appearance (Clarity)	Clear	Clear
Appearance (Colour)	Colourless	Colourless
Appearance (Form)	Liquid	Liquid
Molarity	0.1 ± 0.005	0.1025

FRM/82/11/4/91288 Rev No. 1 Date 08/29/2018

Prepared by **Prajakta**

In compliance with the standard specifications of SRL

Plant Site 2 : H-4, MIDC, Talaja – 410 208, Maharashtra, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

an ISO 9001-2015, 17025:2017, CE, GMP Certified Company

This COA has been generated electronically and it is valid without signature.