



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product	52665 - Chloramphenicol Yeast Glucose Agar BioVeg
Batch No	1306377
Analysis Date	07-May-2024
Date of Manufacture	May 2024
Expiry Date/Re-Test Date	May 2027

Test Parameters	Standards	Actual Results
Appearance (Colour)	Light yellow	Complies
Appearance (Form)	Free flowing, homogeneous powder	Complies
Solubility	40.00 gm/liter	Complies
Solubility before autoclaving (Clarity)	Clear to slightly opalescent	Complies
Gel strength (1.5% gel)	Firm, comparable with 1.5% agar gel	Complies
pH (25°C)	6.6 ± 0.2	6.70 Complies
Prepared Medium Appearance after autoclaving (Clarity)	Clear to very slightly opalescent	Complies
Prepared Medium Appearance after autoclaving (Colour)	Light amber	Complies
Cultural Response	Inoculate and incubate at 25 ± 2°C for 2 - 5 days	
Organism	Candida albicans ATCC 10231	
Inoculum (cfu) 10-100	Growth : Good	Complies
Organism	Aspergillus niger ATCC 16404	
Inoculum	Growth : Good	Complies
Organism	Saccharomyces cerevisiae ATCC 9763	
Inoculum (cfu) 10-100	Growth : Good	Complies
Organism	Escherichia coli ATCC 25922	
Inoculum	Growth : Inhibited	Complies
Organism	Staphylococcus aureus ATCC 25923	
Inoculum	Growth : Inhibited	Complies

FRM/82/11/3/52665 Rev No. 2 Date 26/12/2021

Prepared by Shikha

In compliance with the standard specifications of SRL

Plant Site 1 : D-88/2, MIDC, Turbhe – 400 705, New Mumbai, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.