



SISCO
RESEARCH
LABORATORIES
PVT. LTD.

www.srlchem.com

Certificate of Analysis

Product	57380 - Copper AAS Standard Solution in 1N HNO₃ (Traceable to NIST)
Batch No	1734674
Analysis Date	16-Apr-2024
Date of Manufacture	April 2024
Expiry Date/Re-Test Date	April 2027
Molecular Formula	221.26

Test Parameters	Standards	Actual Results
Appearance (Clarity)	Clear	Clear
Appearance (Colour)	Blue	Blue
Appearance (Form)	Liquid	Liquid
Concentration (Cu)	1000 ± 50 mg/Ltr	1047 mg/Ltr
Traceability	Traceable to NIST	Which is traceable to HC31103786 is Directly traceable to NIST SRM 3114 Lot No. 120618

FRM/82/11/4/57380 Rev No. 2 Date 07/03/2024

Prepared by Vivek

In compliance with the standard specifications of SRL

Plant Site 2 : H-4, MIDC, Talaja – 410 208, Maharashtra, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.