



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product

**27769 - Selenium AAS Standard Solution in 0.5N
HNO₃**

Batch No

1896436

Analysis Date

13-Oct-2022

Date of Manufacture

October 2022

Expiry Date/Re-Test Date

October 2025

Test Parameters	Standards	Actual Results
Appearance (Clarity)	Clear	Clear
Appearance (Colour)	Colourless	Colourless
Appearance (Form)	Liquid	Liquid
Concentration (Se)	1000 ± 50 mg/Ltr	1020 mg/Ltr
Traceability		Which is traceable to Sigma Aldrich Merck Batch No: HC02386596 Product No: 1.19796.0100 Which is traceable to NIST

FRM/82/11/4/27769 Rev No. 1 Date 02/07/2022

Prepared by **Mayur**

In compliance with the standard specifications of SRL

Plant Site 2 : H-4, MIDC, Talaja – 410 208, Maharashtra, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

an ISO 9001-2015, 17025:2017, CE, GMP Certified Company

This COA has been generated electronically and it is valid without signature.