



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product	12447 - Mid-Range DNA Marker: 0.5 – 5 kb
Batch No	1897139
Analysis Date	04-Jan-2023
Date of Manufacture	January 2023
Expiry Date/Re-Test Date	January 2025

Test Parameters	Standards	Actual Results
Appearance (Form)	Liquid	Liquid
Appearance (Colour)	Blue	Blue
Appearance (Clarity)	Clear	Clear
Electrophoresis test	Well-defined bands are formed during agarose gel electrophoresis.	Conforms
Stability	No obvious degradation was seen after incubation at 37°C for 1 week	Conforms
Storage	-20°C	Maintained

FRM/82/11/5/12447 Rev No. 1 Date 18/06/2020

Prepared by **Soumya**

In compliance with the standard specifications of SRL

Plant Site 3 : No. 21, Thomas Nagar, Little Mount, Chennai - 600 015, India.

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

This COA has been generated electronically and it is valid without signature.

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)