



Certificate of Analysis

Product **25844 - Kligler Iron Agar BioVeg**
Batch No **2208991**
Analysis Date 13-Jan-2022
Date of Manufacture January 2022
Expiry Date/Re-Test Date January 2025

Test Parameters	Standards	Actual Results
Appearance (Colour)	Light pink	Complies
Appearance (Form)	Free flowing, homogeneous powder	Complies
Solubility	57.52 gm/liter	Complies
pH (25°C)	7.4 ± 0.2	7.35, Complies
Prepared Medium Appearance after autoclaving (Clarity)	Clear to very slightly opalescent	Complies
Prepared Medium Appearance after autoclaving (Colour)	Red	Complies
Cultural Response	Inoculate and incubate at 37°C for 18 - 24 hours	
Organism	Escherichia coli ATCC 25922	
Inoculum (cfu) 10-100	Growth : Good Slant : A Butt : A Gas : (+)ve H ₂ S : (-) ve	Complies
Organism	Enterobacter aerogenes ATCC 13048	
Inoculum (cfu) 10-100	Growth : Good Slant : A Butt : A Gas : (+)ve H ₂ S : (-) ve Gas : (+) ve	Complies
Organism	Salmonella typhimurium ATCC 14028	
Inoculum (cfu) 10-100	Growth : Good Slant : K Butt : A Gas : (+) ve H ₂ S : (+) ve Gas : (+) ve	Complies
Organism	Klebsiella pneumoniae ATCC 13883	



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product **25844 - Kligler Iron Agar BioVeg**
Batch No **2208991**
Analysis Date 13-Jan-2022
Date of Manufacture January 2022
Expiry Date/Re-Test Date January 2025

Test Parameters	Standards	Actual Results
Inoculum (cfu) 10-100	Growth : Good Slant : A Butt : A Gas : (+)ve H ₂ S : (-) ve	Complies
Organism	<i>Pseudomonas aeruginosa</i> ATCC 27853	
Inoculum (cfu) 10-100	Growth : Good Slant : K Butt : K Gas : (-)ve H ₂ S : (-) ve	Complies
Organism	<i>Proteus vulgaris</i> ATCC 13315	
Inoculum (cfu) 10-100	Growth : Good Slant : K Butt : A Gas : (-) ve H ₂ S : (+) ve	Complies
Organism	<i>Shigella flexneri</i> ATCC 12022	
Inoculum (cfu) 10-100	Growth : Good Slant : K Butt : A Gas : (-) ve H ₂ S : (-) ve	Complies
Organism	<i>Salmonella typhi</i> ATCC 6539	
Inoculum (cfu) 10-100	Growth : Good Slant : K Butt : A Gas : -ve H ₂ S : +ve	Complies
Organism	<i>Salmonella Paratyphi A</i> ATCC 5006	

FRM/82/11/3/25844 Rev No. 1 Date 27/12/2021

Prepared by Shikha

In compliance with the standard specifications of SRL

Plant Site 1 : D-88/2, MIDC, Turbhe – 400 705, New Mumbai, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

2/3

This COA has been generated electronically and it is valid without signature.

[an ISO 9001-2015 Certified Company](#)



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product	25844 - Kligler Iron Agar BioVeg
Batch No	2208991
Analysis Date	13-Jan-2022
Date of Manufacture	January 2022
Expiry Date/Re-Test Date	January 2025

Test Parameters	Standards	Actual Results
Inoculum (cfu) 10-100	Growth : Good Slant : K Butt : A Gas : +ve H2S : -ve	Complies
Organism	Salmonella enteritidis ATCC 13076	
Inoculum (cfu) 10-100	Growth : Good Slant : K Butt : K Gas : (+) ve H2S : (+) ve	Complies
Key	A = acidic, yellow; K = alkaline, no change; Gas = (+) splitting and cracking of the medium, Gas = (-) no splitting and cracking of the medium, H2S = (+) black precipitate; H2S = (-) no black precipitate	

FRM/82/11/3/25844 Rev No. 1 Date 27/12/2021

Prepared by Shikha

In compliance with the standard specifications of SRL

Plant Site 1 : D-88/2, MIDC, Turbhe – 400 705, New Mumbai, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

3/3

This COA has been generated electronically and it is valid without signature.

[an ISO 9001-2015 Certified Company](#)