



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product	11999 - Reinforced Medium for Clostridia, Harmonized
Batch No	2229196
Analysis Date	01-Aug-2022
Date of Manufacture	August 2022
Expiry Date/Re-Test Date	August 2025

Test Parameters	Standards	Actual Results
Appearance (Colour)	Light tan	Complies
Appearance (Form)	Free flowing, homogeneous powder	Complies
Solubility	38.00 gm/liter	Complies
Solubility before autoclaving (Clarity)	Clear to very slightly opalescent solution	Complies
pH (25°C)	6.8 ± 0.2	6.93, Complies
Prepared Medium Appearance after autoclaving (Clarity)	Clear to very slightly opalescent solution	Complies
Prepared Medium Appearance after autoclaving (Colour)	Medium amber	Complies
Cultural Response	Inoculate and incubate at 35 ± 2°C for 18-48 hours	
Organism	Clostridium perfringens ATCC 13124	
Inoculum (cfu) 10-100	Growth : Good	Complies
Organism	Bacteroides vulgatus ATCC 8482	
Inoculum (cfu) 10-100	Growth : Good	Complies
Organism	Bacteroides fragilis ATCC 23745	
Inoculum (cfu) 10-100	Growth : Good	Complies
Organism	Clostridium sporogenes ATCC 13124	
Inoculum (cfu) 10-100	Growth : Good	Complies

FRM/82/11/3/11999 Rev No. 1 Date 09/06/2022

Prepared by Shikha

In compliance with the standard specifications of SRL

Plant Site 1 : D-88/2, MIDC, Turbhe – 400 705, New Mumbai, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

This COA has been generated electronically and it is valid without signature.

[an ISO 9001-2015 Certified Company](#)