



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product	92096 - Perfringens S.F.P. Supplement (S.F.P. Supplement)
Batch No	2277342
Analysis Date	30-Jan-2020
Date of Manufacture	January 2020
Expiry Date/Re-Test Date	January 2022

Test Parameters	Standards	Actual Results
Appearance (Colour)	White, fine powder	Complies
Solubility	Soluble in distilled water	Complies
Cultural Response	Inoculate and incubate at 35 ± 2°C for 18-24 hours under anaerobic conditions (Aseptically, add the rehydrated contents of 1 vial of Perfringens S.F.P. Supplement (92096) as per direction)	
Organism	Clostridium perfringens ATCC 12924	
Inoculum (cfu) 10-100	Growth : Good Sulphite reduction : Positive	Complies
Organism	Escherichia coli ATCC 25922	
Inoculum	Growth : Inhibited	Complies

FRM/82/11/3/92096 Rev No. 1 Date 07/08/2023

Prepared by Shikha

In compliance with the standard specifications of SRL

Plant Site 1 : D-88/2, MIDC, Turbhe – 400 705, New Mumbai, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.