



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product	37351 - RNA Mini Spin Column Set (2ml SC w/ CT)
Batch No	2392940
Analysis Date	01-Sep-2020
Date of Manufacture	September 2020
Expiry Date/Re-Test Date	September 2023

Test Parameters	Standards	Actual Results
Appearance (Form)	Circular Columns and Collection tubes	Conforms
Absence of nucleases	Nucleic acids are incubated in the tube at 37°C for 1 hour	No degradation seen
Absorbance	20-40µg	Conforms
Stability	Stable if stored as recommended	Conforms
Storage	Store in cool dry place	Maintained

FRM/82/11/5/37351 Rev No. 1 Date 30/07/2020

Prepared by **Soumya**

In compliance with the standard specifications of SRL

Plant Site 3 : No. 21, Thomas Nagar, Little Mount, Chennai - 600 015, India.

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099, India. ☎ +91-22-4268 5800

[an ISO 9001-2015 Certified Company](#)

This COA has been generated electronically and it is valid without signature.