



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product	67124 - Nalidixic Selective Supplement (15mg/vl)
Batch No	2691077
Analysis Date	11-Dec-2022
Date of Manufacture	December 2022
Expiry Date/Re-Test Date	December 2024

Test Parameters	Standards	Actual Results
Appearance (Colour)	White, fine powder	Complies
Solubility	Soluble in 0.2 N sodium hydroxide	Complies
Cultural Response	Inoculate and incubate at 35 ± 2°C for 24-48 hours. (Aseptically add the rehydrated contents of 1 vial of Nalidixic Selective Supplement as per direction)	
Organism	Pseudomonas aeruginosa ATCC 27853	
Inoculum (cfu) 10-100	Growth : Good	Complies
Organism	Staphylococcus aureus ATCC 25923	
Inoculum	Growth : Inhibited	Complies
Organism	Escherichia coli ATCC 25922	
Inoculum	Growth : Inhibited	Complies

FRM/82/11/3/67124 Rev No. 1 Date 06/08/2023

Prepared by Shikha

In compliance with the standard specifications of SRL

Plant Site 1 : D-88/2, MIDC, Turbhe – 400 705, New Mumbai, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.