



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product

**64765 - Boron AAS Standard Solution in H2O
(Traceable to NIST)**

Batch No

3395593

Analysis Date

28-Mar-2024

Date of Manufacture

March 2024

Expiry Date/Re-Test Date

March 2027

Test Parameters

Standards

Actual Results

Appearance (Clarity)

Clear

Clear

Appearance (Colour)

Colourless

Colourless

Appearance (Form)

Liquid

Liquid

Concentration (B)

1000 ± 50 mg/Ltr

980 mg/Ltr

Traceability

Traceable to NIST

Which is traceable to NIST
SRM No.3107 Lot 110830;
BAM CRM No: 60357 Lot
BCBP3137V

FRM/82/11/4/64765 Rev No. 2 Date 07/03/2024

Prepared by **Prajakta**

In compliance with the standard specifications of SRL

Plant Site 2 : H-4, MIDC, Talaja – 410 208, Maharashtra, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.