



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product	97409 - Buffer solution pH 7
Batch No	3548114
Analysis Date	19-Mar-2024
Date of Manufacture	March 2024
Expiry Date/Re-Test Date	March 2025

Test Parameters	Standards	Actual Results
Appearance (Clarity)	Clear	Clear
Appearance (Colour)	Colourless	Colourless
Appearance (Form)	Liquid	Liquid
pH (25°C)	6.95 - 7.05	6.99

FRM/82/11/4/97409 Rev No. 1 Date 01/11/2015

Prepared by Prasad

In compliance with the standard specifications of SRL

Plant Site 2 : H-4, MIDC, Talaja – 410 208, Maharashtra, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.