



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product	86183 - Chromium AAS Standard Solution in 0.5N HNO3
Batch No	3734148
Analysis Date	17-Aug-2022
Date of Manufacture	August 2022
Expiry Date/Re-Test Date	August 2025

Test Parameters	Standards	Actual Results
Appearance (Clarity)	Clear	Clear
Appearance (Colour)	Blue	Blue
Appearance (Form)	Liquid	Liquid
Concentration (Cr)	1000 ± 50 mg/Ltr	1040 mg/Ltr
Traceability		This sample is traceable to SRL Batch No: 3205929 Chromium AAS Std is traceable to merck Lot no: 1.19871.052 which is traceable to NIST SRM 119871053A

FRM/82/11/4/86183 Rev No. 1 Date 02/07/2022

Prepared by **Prajakta**

In compliance with the standard specifications of SRL

Plant Site 2 : H-4, MIDC, Talaja – 410 208, Maharashtra, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.