



**SISCO  
RESEARCH  
LABORATORIES  
PVT. LTD.**

[www.srlchem.com](http://www.srlchem.com)

## Certificate of Analysis

<b>Product</b>	<b>69263 - Palladium AAS Standard Solution in 0.5N HCl</b>
<b>Batch No</b>	<b>3888787</b>
<b>Analysis Date</b>	14-Jan-2023
<b>Date of Manufacture</b>	January 2023
<b>Expiry Date/Re-Test Date</b>	January 2026

Test Parameters	Standards	Actual Results
Appearance (Clarity)	Clear	Clear
Appearance (Colour)	Brown	Brown
Appearance (Form)	Liquid	Liquid
Concentration (Pd)	1000 ± 50 mg/Ltr	981.92 mg/Ltr
Traceability		which is traceable to Sigma Aldrich Merck NIST traceable Std : Batch No. : HC99037182, Product no. : 1.14282.0100

FRM/82/11/4/69263 Rev No. 1 Date 02/07/2022

Prepared by **Pranita**

In compliance with the standard specifications of SRL

Plant Site 2 : H-4, MIDC, Taloja – 410 208, Maharashtra, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,  
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

an ISO 9001-2015, 17025:2017, CE, GMP Certified Company

This COA has been generated electronically and it is valid without signature.