



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product	34123 - VCN Supplement
Batch No	3910779
Analysis Date	08-Feb-2023
Date of Manufacture	February 2023
Expiry Date/Re-Test Date	February 2025

Test Parameters	Standards	Actual Results
Appearance (Colour)	White, fine powder	Complies
Solubility	Soluble in distilled water	Complies
Cultural Response	Inoculate and incubate at 35 ± 2°C for 18-48 hours (Aseptically add the contents of one vial of VCN Supplement (34123) as per direction)	
Organism	Neisseria meningitidis ATCC 13090	
Inoculum (cfu) 10-100	Growth : Good Colony colour : Medium to large, blue-gray, mucoid	Complies
Organism	Escherichia coli ATCC 25922	
Inoculum	Growth : Inhibited	Complies
Organism	Proteus mirabilis ATCC 25933	
Inoculum	Growth : Inhibited	Complies

FRM/82/11/3/34123 Rev No. 2 Date 08/08/2023

Prepared by Shikha

In compliance with the standard specifications of SRL

Plant Site 1 : D-88/2, MIDC, Turbhe – 400 705, New Mumbai, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.