



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product **51747 - Bacillus Medium**
Batch No **4170139**
Analysis Date 01-Oct-2022
Date of Manufacture October 2022
Expiry Date/Re-Test Date October 2025

Test Parameters	Standards	Actual Results
Appearance (Colour)	White to Cream	Complies
Appearance (Form)	Free flowing, homogeneous powder	Complies
Solubility	7.50 gm/liter	Complies
Solubility before autoclaving (Clarity)	Clear solution with very slight precipitate	Complies
pH (25°C)	7.4 ± 0.2	7.21, Complies
Prepared Medium Appearance after autoclaving (Clarity)	Clear solution with very slight precipitate	Complies
Prepared Medium Appearance after autoclaving (Colour)	Cream to yellow	Complies
Cultural Response	Inoculate and incubate at 35 - 37°C for 18- 24 hours	
Organism	Bacillus cereus ATCC 10876	
Inoculum (cfu) 10-100	Growth : Good	Complies
Organism	Bacillus licheniformis ATCC 9945	
Inoculum (cfu) 10-100	Growth : Good	Complies

FRM/82/11/3/51747 Rev No. 1 Date 27/07/2021

Prepared by Pooja

In compliance with the standard specifications of SRL

Plant Site 1 : D-88/2, MIDC, Turbhe – 400 705, New Mumbai, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

This COA has been generated electronically and it is valid without signature.

[an ISO 9001-2015 Certified Company](#)