



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product

**36920 - Calcium AAS Standard Solution in 0.5N HNO₃
(Traceable to NIST)**

Batch No

4247175

Analysis Date

06-Jun-2024

Date of Manufacture

June 2024

Expiry Date/Re-Test Date

June 2027

Test Parameters	Standards	Actual Results
Appearance (Clarity)	Clear	Clear
Appearance (Colour)	Colourless	Colourless
Appearance (Form)	Liquid	Liquid
Concentration (Ca)	1000 ± 50 mg/Ltr	1042 mg/Ltr
Traceability	Traceable to NIST	Which is traceable to NIST SRM 3109a, Lot No. 130213

FRM/82/11/4/36920 Rev No. 2 Date 07/03/2024

Prepared by Yogesh

In compliance with the standard specifications of SRL

Plant Site 2 : H-4, MIDC, Talaja – 410 208, Maharashtra, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

This COA has been generated electronically and it is valid without signature.

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)