



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product	75105 - Lithium AAS Standard Solution in 0.5N HNO₃
Batch No	4406894
Analysis Date	04-Dec-2023
Date of Manufacture	December 2023
Expiry Date/Re-Test Date	December 2026

Test Parameters	Standards	Actual Results
Appearance (Clarity)	Clear	Clear
Appearance (Colour)	Colourless	Colourless
Appearance (Form)	Liquid	Liquid
Concentration (Li)	1000 ± 50 mg/Ltr	1012 mg/Ltr
Traceability		Which is traceable to Sigma Aldrich Merck product no :- 1.70223.0100 Lot no :- HC17013123 Which is traceable to NIST SRM 3129a, Lot 100714.

FRM/82/11/4/75105 Rev No. 1 Date 02/07/2022

Prepared by Vivek

In compliance with the standard specifications of SRL

Plant Site 2 : H-4, MIDC, Talaja – 410 208, Maharashtra, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.