



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product	75063 - Sodium AAS Standard Solution in 0.5N HCl
Batch No	4510397
Analysis Date	19-Oct-2022
Date of Manufacture	October 2022
Expiry Date/Re-Test Date	October 2025

Test Parameters	Standards	Actual Results
Appearance (Clarity)	Clear	Clear
Appearance (Colour)	Colourless	Colourless
Appearance (Form)	Liquid	Liquid
Concentration (Na)	1000 ± 50 mg/Ltr	988 mg/Ltr
Traceability		Which is traceable to SRL Batch No: 7111875 Which is traceable to Thomas baker Lot No: 743380 Which is traceable to NIST

FRM/82/11/4/75063 Rev No. 2 Date 02/07/2022

Prepared by **Mayur**

In compliance with the standard specifications of SRL

Plant Site 2 : H-4, MIDC, Talaja – 410 208, Maharashtra, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

an ISO 9001-2015, 17025:2017, CE, GMP Certified Company

This COA has been generated electronically and it is valid without signature.