



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product

**92035 - 40% Acrylamide / Bis-acrylamide Mix Solution
(Ratio: 37.5:1)**

Batch No

4665793

Analysis Date

31-Jan-2024

Date of Manufacture

January 2024

Expiry Date/Re-Test Date

January 2026

Test Parameters

Standards

Actual Results

Appearance (Clarity)

Clear

Clear

Appearance (Colour)

Colourless

Colourless

Appearance (Form)

Liquid

Liquid

pH

5.0 - 8.0

7.6

Specific conductance (25°C)

max. 10 µmho/cm

8.18 µmho/cm

Suitability for use in gel

Passes Test

Passes Test

Electrophoresis

FRM/82/11/4/92035 Rev No. 1 Date 29/07/2023

Prepared by **Mayur**

In compliance with the standard specifications of SRL

Plant Site 2 : H-4, MIDC, Talaja – 410 208, Maharashtra, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.