



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product	93834 - Clostridium Difficile Supplement
Batch No	5385952
Analysis Date	20-Aug-2023
Date of Manufacture	August 2023
Expiry Date/Re-Test Date	August 2025

Test Parameters	Standards	Actual Results
Appearance (Colour)	White, fine powder.	Complies
Solubility	Soluble in distilled water.	Complies
Cultural Response	Inoculate and incubate at 35 ± 2°C for 48 hours. (Aseptically add the rehydrated contents of 1 vial as per direction) Growth observed on Clostridium Difficile Agar Base.	
Organism	Clostridium difficile ATCC 11204	
Inoculum (cfu) 10-100	Growth : Good. Colony colour : Greyish white.	Complies
Organism	Shigella flexneri ATCC 12022	
Inoculum	Growth : Inhibited	Complies
Organism	Escherichia coli ATCC 25922	
Inoculum	Growth : Inhibited	Complies
Organism	Staphylococcus aureus ATCC 25923	
Inoculum	Growth : Inhibited	Complies

FRM/82/11/3/93834 Rev No. 1 Date 03/08/2023

Prepared by Shikha

In compliance with the standard specifications of SRL

Plant Site 1 : D-88/2, MIDC, Turbhe – 400 705, New Mumbai, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

This COA has been generated electronically and it is valid without signature.

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)