



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product **55582 - Bolton Selective Supplement**
Batch No **5409298**
Analysis Date 24-Mar-2022
Date of Manufacture March 2022
Expiry Date/Re-Test Date March 2024

| Test Parameters | Standards | Actual Results |
|-----------------------|--|----------------|
| Appearance (Colour) | White, fine powder | Complies |
| Solubility | Completely soluble in 50.0% v/v aqueous ethanol | Complies |
| Cultural Response | Inoculate and incubate at 35 ± 2°C for 4-6 hours and then at 41.5°C for 40 - 48 hours. (Aseptically add the contents of one vial as per direction) | |
| Organism | Campylobacter coli ATCC 3359 | |
| Inoculum (cfu) 10-100 | Growth : Good | Complies |
| Organism | Campylobacter jejuni ATCC 29428 | |
| Inoculum (cfu) 10-100 | Growth : Good | Complies |
| Organism | Escherichia coli ATCC 25922 | |
| Inoculum | Growth : Inhibited | Complies |
| Organism | Candida albicans ATCC 10231 | |
| Inoculum | Growth : Inhibited | Complies |

FRM/82/11/3/55582 Rev No. 2 Date 11/04/2024

Prepared by Shikha

In compliance with the standard specifications of SRL

Plant Site 1 : D-88/2, MIDC, Turbhe – 400 705, New Mumbai, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.