



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product	29760 - Klebsiella Selective Supplement (25mg/vl)
Batch No	5601121
Analysis Date	07-Feb-2022
Date of Manufacture	February 2022
Expiry Date/Re-Test Date	February 2024

Test Parameters	Standards	Actual Results
Appearance (Colour)	White, fine powder	Complies
Solubility	Soluble in distilled water	Complies
Cultural Response	Inoculate and incubate at 35 ± 2°C for 18-24 hours. (Aseptically add the rehydrated contents of 1 vial of Klebsiella Selective Supplement as per direction)	
Organism	Klebsiella pneumoniae ATCC 13883	
Inoculum (cfu) 10-100	Growth : Good Colony colour : Purple-magenta (mucoid)	Complies
Organism	Escherichia coli ATCC 25922	
Inoculum	Growth : Inhibited	Complies
Organism	Enterobacter aerogenes ATCC 13048	
Inoculum	Growth : Inhibited	Complies

FRM/82/11/3/29760 Rev No. 1 Date 06/08/2023

Prepared by Shikha

In compliance with the standard specifications of SRL

Plant Site 1 : D-88/2, MIDC, Turbhe – 400 705, New Mumbai, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.