

Certificate of Analysis

Product

52071 - Fraser Selective Supplement

Batch No Analysis Date

5733153 12-Jan-2023 January 2023

Date of Manufacture
Expiry Date/Re-Test Date

January 2025

Tool Dovementous	Ctorodoudo	A atual Daguita
Test Parameters	Standards	Actual Results
Cultural Response	Inoculate and incubate at 35 ± 2°C for 24-48 hours (Aseptically add the contents of one vial of Fraser Selective Supplement (52071))	
Organism	Listeria monocytogenes ATCC 19117	
Inoculum (cfu) 10-100	Growth : Good, Reaction : + (positive, blackening of medium)	Complies
Organism	Pseudomonas aeruginosa ATCC 27853	
Inoculum (cfu) 10-100	Growth : Good, Reaction : - (negative, no blackening of medium)	Complies

FRM/82/11/3/52071 Rev No. 1 Date 13/12/2021

Prepared by Shikha

In compliance with the standard specifications of SRL