



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product	42116 - Oxford Listeria Supplement
Batch No	6416609
Analysis Date	05-Dec-2023
Date of Manufacture	December 2023
Expiry Date/Re-Test Date	December 2025

Test Parameters	Standards	Actual Results
Appearance (Colour)	White to Off-White powder	Complies
Solubility	Completely soluble in 50.0% Ethanol	Complies
Cultural Response	Inoculate and incubate at $35 \pm 2^{\circ}\text{C}$ for 24-48 hours (Aseptically, add the rehydrated contents of 1 vial of Oxford Listeria Supplement (42116))	
Organism	Listeria monocytogenes ATCC 19112	
Inoculum (cfu) 10-100	Growth : Good Esculin hydrolysis : Positive reaction, Blackening of medium around colony	Complies
Organism	Staphylococcus aureus ATCC 25923	
Inoculum (cfu) 10-100	Growth : Good Esculin hydrolysis : Negative reaction	Complies
Organism	Escherichia coli ATCC 25922	
Inoculum	Growth : Inhibited	Complies

FRM/82/11/3/42116 Rev No. 2 Date 15/12/2021

Prepared by Shikha

In compliance with the standard specifications of SRL

Plant Site 1 : D-88/2, MIDC, Turbhe – 400 705, New Mumbai, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.