



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product	92605 - Water for HPLC - [7732-18-5]
Batch No	6426675
Analysis Date	18-Feb-2024
Date of Manufacture	February 2024
Expiry Date/Re-Test Date	February 2025
Molecular Formula	H2O
Molecular Weight	18.02

Test Parameters	Standards	Actual Results
Appearance (Clarity)	Clear	Clear
Appearance (Colour)	Colourless	Colourless
Appearance (Form)	Liquid	Liquid
Chloride (Cl)	max 0.000001%	Passes
Sulphate (SO ₄)	max 0.00001%	Passes
Non Volatile Matter	max 0.0005%	0.0001%
Absorbance of sample in 1cm cell v/s H ₂ O		
@202nm	max. 0.01	0.002
@210nm	max. 0.01	0.001
@250 nm	max. 0.005	0.001
@400nm	max. 0.005	0.002
Suitability for HPLC	Passes test	Passes

FRM/82/11/4/92605 Rev No. 1 Date 01/11/2015

Prepared by **Prajakta**

In compliance with the standard specifications of SRL

Plant Site 2 : H-4, MIDC, Taloja – 410 208, Maharashtra, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.