



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product	90508 - 10x Roseate PCR Buffer with Mg2+
Batch No	6519089
Analysis Date	19-Jan-2023
Date of Manufacture	January 2023
Expiry Date/Re-Test Date	January 2025

Test Parameters	Standards	Actual Results
Appearance (Form)	Liquid	Liquid
Appearance (Colour)	Dark Pink	Dark Pink
Appearance (Clarity)	Clear	Clear
PCR amplification	Amplification using the buffer in the reaction mix for PCR should give the required amplicons	Conforms
Absence of nucleases	Incubation at 37°C of the product with plasmid DNA for 24 hours should show no degradation of DNA.	Conforms
Storage	-20°C	Maintained

FRM/82/11/5/90508 Rev No. 1 Date 05/10/2020

Prepared by KalaiSelvi

In compliance with the standard specifications of SRL

Plant Site 3 : No. 21, Thomas Nagar, Little Mount, Chennai - 600 015, India.

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

This COA has been generated electronically and it is valid without signature.

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)