



SISCO
RESEARCH
LABORATORIES
PVT. LTD.

www.srlchem.com

Certificate of Analysis

Product 94486 - Manganese AAS Standard Solution in 0.5N HCl
(Traceable to NIST)

Batch No 6680268
Analysis Date 29-Apr-2024
Date of Manufacture April 2024
Expiry Date/Re-Test Date April 2027

Test Parameters	Standards	Actual Results
Appearance (Clarity)	Clear	Clear
Appearance (Colour)	Colourless	Colourless
Appearance (Form)	Liquid	Liquid
Concentration (Mn)	1000 ± 50 mg/ Ltr	1046 mg/ Ltr
Traceability	Traceable to NIST	Which is traceable to NIST batch No. HC04409084 Lot No. 1.19789.0100

FRM/82/11/4/94486 Rev No. 2 Date 07/03/2024

Prepared by Dhanashree

In compliance with the standard specifications of SRL

Plant Site 2 : H-4, MIDC, Talaja – 410 208, Maharashtra, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.