



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product 91488 - Potassium Tellurite 1% Supplement (1ml/vl)
Batch No 6823066
Analysis Date 04-Feb-2024
Date of Manufacture February 2024
Expiry Date/Re-Test Date February 2026

Test Parameters	Standards	Actual Results
Appearance (Colour)	Clear solution.	Complies
Solubility	1 %, Clear solution.	Complies
Cultural Response	Inoculate and incubate at 35 ± 2°C for 18-48 hours. (Aseptically add the contents of 1 vial as per direction) Growth observed on Baird-Parker Agar, Base (B/S)(24562).	
Organism	Staphylococcus aureus ATCC 25923	
Inoculum (cfu) 10-100	Growth : Good Colony colour : Grey-black, shiny. Lecithinase halo : Positive.	Complies
Organism	Staphylococcus epidermidis ATCC 12228	
Inoculum (cfu) 10-100	Growth : Poor to good. Colony colour : Black. Lecithinase halo : Negative.	Complies
Organism	Bacillus subtilis ATCC 6633	
Inoculum (cfu) 10-100	Growth : None to poor. Colony colour : Brown. Lecithinase halo : Negative.	Complies
Organism	Proteus mirabilis ATCC 25933	
Inoculum (cfu) 10-100	Growth : Good. Colony colour : Brown-black. Lecithinase halo : Negative.	Complies

FRM/82/11/3/91488 Rev No. 1 Date 03/08/2023

Prepared by Shikha

In compliance with the standard specifications of SRL

Plant Site 1 : D-88/2, MIDC, Turbhe – 400 705, New Mumbai, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.