

Certificate of Analysis

Product

74183 - Brilliant Yellow Indicator Papers

Batch No
Analysis Date

6920567

Date of Manufacture

06-Sep-2022 September 2022

Expiry Date/Re-Test Date

September 2027

Test Parameters	Standards	Actual Results
Suitability as pH indicator	Passes test (pH 6.7 Yellow pH 7.9	Passes
	Red)	

FRM/82/11/4/74183 Rev No. 1 Date 01/11/2015

Prepared by Dnyaneshwar

In compliance with the standard specifications of SRL

Plant Site 2 : H-4, MIDC, Taloja – 410 208, Maharashtra, India