



SISCO
RESEARCH
LABORATORIES
PVT. LTD.

www.srlchem.com

Certificate of Analysis

Product

**32737 - Silver AAS Standard Solution in 0.5N HNO3
(Traceable to NIST)**

Batch No

6952034

Analysis Date

06-Mar-2024

Date of Manufacture

March 2024

Expiry Date/Re-Test Date

March 2027

Test Parameters

Standards

Actual Results

Appearance (Clarity)

Clear

Clear

Appearance (Colour)

Colourless

Colourless

Appearance (Form)

Liquid

Liquid

Concentration (Ag)

1000 ± 50 mg/Ltr

1012 mg/Ltr

Traceability

Traceable to NIST

Whitch is traceable to sigma
aldrich Merck NIST traceable
std Batch no. HC03474497
Product No. 1.19797.0100

FRM/82/11/4/32737 Rev No. 2 Date 07/03/2024

Prepared by **Prajakta**

In compliance with the standard specifications of SRL

Plant Site 2 : H-4, MIDC, Talaja – 410 208, Maharashtra, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.