



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product **51993 - Soyabean Casein Digest Medium (SCDM)
(Tryptone Soya Broth, TSB) BioVeg**

Batch No **7127835**
Analysis Date 02-Jul-2024
Date of Manufacture July 2024
Expiry Date/Re-Test Date July 2027

Test Parameters	Standards	Actual Results
Appearance (Colour)	Light cream	Complies
Appearance (Form)	Free flowing, homogeneous powder	Complies
Solubility	30.00 gm/liter	Complies
Solubility before autoclaving (Clarity)	Clear without any precipitate	Complies
pH (25°C)	7.3 ± 0.2	7.43 Complies
Prepared Medium Appearance after autoclaving (Clarity)	Clear without any precipitate	Complies
Prepared Medium Appearance after autoclaving (Colour)	Light amber	Complies
Cultural Response	Inoculate and incubate at 37°C for 18 - 24 hours for bacteria and at 25°C for fungi* for 2 - 7 days.	
Organism	Candida albicans * ATCC 10231	
Inoculum (cfu) 10-100	Growth : Good	Complies
Organism	Staphylococcus aureus ATCC 25923	
Inoculum (cfu) 10-100	Growth : Good	Complies
Organism	Streptococcus pyogenes ATCC 19615	
Inoculum (cfu) 10-100	Growth : Good	Complies
Organism	Bacillus subtilis ATCC 6633	
Inoculum (cfu) 10-100	Growth : Good	Complies
Organism	Bacteroides vulgatus ATCC 8482	
Inoculum (cfu) 10-100	Growth : Good	Complies
Organism	Escherichia coli ATCC 25922	
Inoculum (cfu) 10-100	Growth : Good	Complies
Organism	Pseudomonas aeruginosa ATCC 27853	
Inoculum (cfu) 10-100	Growth : Good	Complies

FRM/82/11/3/51993 Rev No. 1 Date 19/12/2021

Prepared by **Samrudhi**

In compliance with the standard specifications of SRL

Plant Site 1 : D-88/2, MIDC, Turbhe – 400 705, New Mumbai, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/2

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](https://www.iso.org/standard/68549.html)

This COA has been generated electronically and it is valid without signature.



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product

**51993 - Soyabean Casein Digest Medium (SCDM)
(Tryptone Soya Broth, TSB) BioVeg**

Batch No

7127835

Analysis Date

02-Jul-2024

Date of Manufacture

July 2024

Expiry Date/Re-Test Date

July 2027

Test Parameters	Standards	Actual Results
Inoculum (cfu) 10-100	Micrococcus luteus ATCC 9341	
Organism	Growth : Good	Complies

FRM/82/11/3/51993 Rev No. 1 Date 19/12/2021

Prepared by **Samrudhi**

In compliance with the standard specifications of SRL

Plant Site 1 : D-88/2, MIDC, Turbhe – 400 705, New Mumbai, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

2/2

This COA has been generated electronically and it is valid without signature.

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)