

Certificate of Analysis

Product

48234 - ProxiO DNA Marker 2: 0.1-1.2kb

Batch No Analysis Date Date of Manufacture **7168297** 24-Jan-2023 January 2023

Expiry Date/Re-Test Date

January 2025

Test Parameters	Standards	Actual Results
Appearance (Form)	Liquid	Liquid
Appearance (Colour)	Blue	Blue
Appearance (Clarity)	Clear	Clear
Agarose gel electrophoresis	Well defined bands seen	Conforms
A260/A280	1.7~1.9	Conforms
Stability		Conforms
	No degradation seen after incubation at	
	37°C for one week	
Storage	-20°C	Maintained

FRM/82/11/5/48234 Rev No. 1 Date 16/09/2020

Prepared by KalaiSelvi

In compliance with the standard specifications of SRL