



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product	50900 - Taq Mix (2x) (PCR Master Mix (2x))
Batch No	7411873
Analysis Date	26-Mar-2024
Date of Manufacture	March 2024
Expiry Date/Re-Test Date	March 2026

Test Parameters	Standards	Actual Results
Appearance (Form)	Liquid	Liquid
Appearance (Colour)	Blue	Blue
Appearance (Clarity)	Clear	Clear
Activity	Good and pure yield of PCR product that can be used for downstream process	Conforms
Amplification capability	Upto 5kb DNA fragment	Conforms
Absence of nucleases	No DNA degradation seen when incubated overnight with the Master mix.	Conforms
DNA contamination	No Plasmid or Genomic DNA contamination seen	Conforms
Storage	-20°C	Maintained

FRM/82/11/5/50900 Rev No. 1 Date 04/03/2020

Prepared by KalaiSelvi

In compliance with the standard specifications of SRL

Plant Site 3 : No. 21, Thomas Nagar, Little Mount, Chennai - 600 015, India.

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.