



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product 31199 - Cetrime Broth (U/P)
Batch No 7937547
Analysis Date 06-Dec-2023
Date of Manufacture December 2023
Expiry Date/Re-Test Date December 2026

Test Parameters	Standards	Actual Results
Appearance (Colour)	Yellow	Complies
Appearance (Form)	Free flowing, homogeneous powder	Complies
Solubility	25.30 gm/liter	Complies
Solubility before autoclaving (Clarity)	Clear to very slightly opalescent without any precipitate	Complies
pH (25°C)	7.2 ± 0.2	7.18 Complies
Prepared Medium Appearance after autoclaving (Clarity)	Clear to very slightly opalescent without any precipitate	Complies
Prepared Medium Appearance after autoclaving (Colour)	Yellow	Complies
Cultural Response	Inoculate and incubate at 35°C for upto 24-48 hours	
Organism	Pseudomonas aeruginosa ATCC 27853	
Inoculum (cfu) 10-100	Growth : Good	Complies
Organism	Escherichia coli ATCC 25922	
Inoculum	Growth : Inhibited	Complies
Organism	Staphylococcus aureus ATCC 25923	
Inoculum	Growth : Inhibited	Complies

FRM/82/11/3/31199 Rev No. 1 Date 05/12/2021

Prepared by Shikha

In compliance with the standard specifications of SRL

Plant Site 1 : D-88/2, MIDC, Turbhe – 400 705, New Mumbai, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.