



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product	56009 - Barium AAS Standard Solution in 0.5N HNO3
Batch No	8236951
Analysis Date	11-Dec-2022
Date of Manufacture	December 2022
Expiry Date/Re-Test Date	December 2025

Test Parameters	Standards	Actual Results
Appearance (Clarity)	Clear	Clear
Appearance (Colour)	Colourless	Colourless
Appearance (Form)	Liquid	Liquid
Concentration (Br)	1000 ± 50 mg/Ltr	1025 mg/Ltr
Traceability		which is traceable to NIST CRM used : MES 3: ICP Alkali's & Alkaline earths

FRM/82/11/4/56009 Rev No. 1 Date 02/07/2022

Prepared by Yogesh

In compliance with the standard specifications of SRL

Plant Site 2 : H-4, MIDC, Talaja – 410 208, Maharashtra, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.