



**SISCO  
RESEARCH  
LABORATORIES  
PVT. LTD.**

[www.srlchem.com](http://www.srlchem.com)

## Certificate of Analysis

<b>Product</b>	<b>41212 - 10X PCR Buffer with 15mM Mg2+</b>
<b>Batch No</b>	<b>8240024</b>
<b>Analysis Date</b>	08-Nov-2022
<b>Date of Manufacture</b>	November 2022
<b>Expiry Date/Re-Test Date</b>	November 2024

Test Parameters	Standards	Actual Results
Appearance (Form)	Liquid	Liquid
Appearance (Colour)	Colourless	Colourless
Appearance (Clarity)	Clear	Clear
PCR amplification	Amplification of standard template along with Taq DNA Polymerase, dNTP, corresponding primers should result in expected molecular size amplicon	Conforms
Absence of nucleases	Incubation at 37°C of the product with Plasmid DNA for 24 hours should show no degradation of the DNA	Conforms
Storage	-20°C	Maintained

FRM/82/11/5/41212 Rev No. 1 Date 07/03/2020

Prepared by KalaiSelvi

In compliance with the standard specifications of SRL

Plant Site 3 : No. 21, Thomas Nagar, Little Mount, Chennai - 600 015, India.

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,  
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

This COA has been generated electronically and it is valid without signature.

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)