



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product	79326 - Shrimp Alkaline Phosphatase (rSAP) (1U/ul)
Batch No	8501069
Analysis Date	14-Aug-2023
Date of Manufacture	August 2023
Expiry Date/Re-Test Date	August 2025

Test Parameters	Standards	Actual Results
Appearance (Form)	Liquid	Liquid
Appearance (Colour)	Colourless	Colourless
Appearance (Clarity)	Clear	Clear
Assay	Plasmid DNA digested with Restriction enzyme is treated with Ligase enzyme to check if dephosphorylation prevented Ligation.	No Ligation seen. Dephosphorylation successful
Absence of nucleases	Incubated with super-coiled Plasmid DNA, ds DNA ss DNA and RNA samples at 37°C	No degradation seen. No Nucleases present
Storage	-20°C	Maintained

FRM/82/11/5/79326 Rev No. 1 Date 06/11/2020

Prepared by KalaiSelvi

In compliance with the standard specifications of SRL

Plant Site 3 : No. 21, Thomas Nagar, Little Mount, Chennai - 600 015, India.

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.