



**SISCO  
RESEARCH  
LABORATORIES  
PVT. LTD.**

[www.srlchem.com](http://www.srlchem.com)

## Certificate of Analysis

**Product** **11947 - Tryptone Broth BioVeg (Tryptone Water BioVeg)**  
**Batch No** **8895153**  
**Analysis Date** 02-Nov-2022  
**Date of Manufacture** November 2022  
**Expiry Date/Re-Test Date** November 2025

Test Parameters	Standards	Actual Results
Appearance (Colour)	Light cream	Complies
Appearance (Form)	Free flowing, homogeneous powder	Complies
Solubility	15.00 gm/liter	Complies
Solubility before autoclaving (Clarity)	Clear without any precipitate	Complies
pH (25°C)	7.5 ± 0.2	7.37, Complies
Prepared Medium Appearance after autoclaving (Clarity)	Clear without any precipitate	Complies
Prepared Medium Appearance after autoclaving (Colour)	Light amber	Complies
Cultural Response	Inoculate and incubate at 37°C for 18 - 24 hours	
Organism	Escherichia coli ATCC 25922	
Inoculum (cfu) 10-100	Growth : Good Indole production : (+)	Complies
Organism	Proteus vulgaris ATCC 13315	
Inoculum (cfu) 10-100	Growth : Good Indole production : (+)	Complies
Organism	Enterobacter aerogenes ATCC 13048	
Inoculum (cfu) 10-100	Growth : Good Indole production : (-)	Complies
Key	(+) : Positive, presence of pink-red ring at top layer (-) : Negative, no presence of pink-red ring at top layer	

FRM/82/11/3/11947 Rev No. 2 Date 26/12/2021

Prepared by Pooja

In compliance with the standard specifications of SRL

Plant Site 1 : D-88/2, MIDC, Turbhe – 400 705, New Mumbai, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,  
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.