



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product **58498 - MES9: ICP Comprehensive Mix A**
Batch No **9611872**
Analysis Date 20-Jan-2024
Date of Manufacture January 2024
Expiry Date/Re-Test Date January 2027

Test Parameters	Standards	Actual Results
Appearance (Clarity)	Clear	Clear
Appearance (Colour)	Colourless to Pale yellow	Colourless
Appearance (Form)	Liquid	Liquid
Silver (Ag)	10.00 ± 0.05 µg/mL	Passes
Aluminium (Al)	10.00 ± 0.05 µg/mL	Passes
Arsenic (As)	10.00 ± 0.05 µg/mL	Passes
Barium (Ba)	10.00 ± 0.05 µg/mL	Passes
Bismuth (Bi)	10.00 ± 0.05 µg/mL	Passes
Calcium (Ca)	10.00 ± 0.05 µg/mL	Passes
Cadmium (Cd)	10.00 ± 0.05 µg/mL	Passes
Cerium (Ce)	10.00 ± 0.05 µg/mL	Passes
Dysprosium (Dy)	10.00 ± 0.05 µg/mL	Passes
Erbium (Er)	10.00 ± 0.05 µg/mL	Passes
Europium (Eu)	10.00 ± 0.05 µg/mL	Passes
Gallium (Ga)	10.00 ± 0.05 µg/mL	Passes
Gadolinium (Gd)	10.00 ± 0.05 µg/mL	Passes
Holmium (Ho)	10.00 ± 0.05 µg/mL	Passes
Lanthanum (La)	10.00 ± 0.05 µg/mL	Passes
Lutetium (Lu)	10.00 ± 0.05 µg/mL	Passes
Magnesium (Mg)	10.00 ± 0.05 µg/mL	Passes
Sodium (Na)	10.00 ± 0.05 µg/mL	Passes
Neodymium (Nd)	10.00 ± 0.05 µg/mL	Passes
Phosphorus (P)	10.00 ± 0.05 µg/mL	Passes
Lead (Pb)	10.00 ± 0.05 µg/mL	Passes
Praseodymium (Pr)	10.00 ± 0.05 µg/mL	Passes
Rubidium (Rb)	10.00 ± 0.05 µg/mL	Passes
Scandium (Sc)	10.00 ± 0.05 µg/mL	Passes
Selenium (Se)	10.00 ± 0.05 µg/mL	Passes
Samarium (Sm)	10.00 ± 0.05 µg/mL	Passes

FRM/82/11/4/58498 Rev No. 3 Date 24/07/2023

Prepared by Yogesh

In compliance with the standard specifications of SRL

Plant Site 2 : H-4, MIDC, Taloja – 410 208, Maharashtra, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/2

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product 58498 - MES9: ICP Comprehensive Mix A
Batch No 9611872
Analysis Date 20-Jan-2024
Date of Manufacture January 2024
Expiry Date/Re-Test Date January 2027

Test Parameters	Standards	Actual Results
Strontium (Sr)	10.00 ± 0.05 µg/mL	Passes
Terbium (Tb)	10.00 ± 0.05 µg/mL	Passes
Thorium (Th)	10.00 ± 0.05 µg/mL	Passes
Thallium (Tl)	10.00 ± 0.05 µg/mL	Passes
Thulium (Tm)	10.00 ± 0.05 µg/mL	Passes
Uranium (U)	10.00 ± 0.05 µg/mL	Passes
Yttrium (Y)	10.00 ± 0.05 µg/mL	Passes

FRM/82/11/4/58498 Rev No. 3 Date 24/07/2023

Prepared by Yogesh

In compliance with the standard specifications of SRL

Plant Site 2 : H-4, MIDC, Talaja – 410 208, Maharashtra, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

2/2

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.