



**SISCO  
RESEARCH  
LABORATORIES  
PVT. LTD.**

[www.srlchem.com](http://www.srlchem.com)

# Certificate of Analysis

**Product**

**38151 - BioLit FluroGreen qPCR Master Mix (High CAR)  
(2X)**

**Batch No**

**9715231**

**Analysis Date**

15-Mar-2024

**Date of Manufacture**

March 2024

**Expiry Date/Re-Test Date**

March 2026

**Test Parameters**

**Standards**

**Actual Results**

Appearance (Form)

Liquid

Liquid

Appearance (Colour)

Colourless

Colourless

Appearance (Clarity)

Clear

Clear

Cycle Threshold

At least 3 of 4 NTC reactions must yield Ct values  $\geq 35$ . Average Ct for 10 copies of human genomic DNA must be  $\leq 35$

Conforms

Standard curve

The correlation coefficient for the standard curve must be  $\geq 0.99$ . The slope of the standard curve must be between -3.2 and - 3.8

Conforms

Absence of nucleases

No DNA degradation seen when incubated overnight with the Master mix.

Conforms

DNA contamination

No DNA contamination seen.

Conforms

Storage

-20°C

Maintained

FRM/82/11/5/38151 Rev No. 1 Date 06/10/2020

Prepared by KalaiSelvi

In compliance with the standard specifications of SRL

Plant Site 3 : No. 21, Thomas Nagar, Little Mount, Chennai - 600 015, India.

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,  
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.