



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product 18263 - MES7: ICP Common Elements Mix 1
Batch No 1179947
Analysis Date 08-Aug-2023
Date of Manufacture August 2023
Expiry Date/Re-Test Date August 2028

Test Parameters	Standards	Actual Results
Appearance (Clarity)	Clear	Clear
Appearance (Colour)	Colourless to Pale yellow	Colourless
Appearance (Form)	Liquid	Liquid
Cadmium (Cd)	100.0 ± 0.5 µg/mL	Cd 100.0 ± 0.5 µg/mL
Cobalt (Co)	100.0 ± 0.5 µg/mL	Co 100.0 ± 0.5 µg/mL
Chromium (Cr)	100.0 ± 0.5 µg/mL	Cr 100.0 ± 0.5 µg/mL
Copper (Cu)	100.0 ± 0.5 µg/mL	Cu 99.99 ± 0.5 µg/mL
Iron (Fe)	100.0 ± 0.5 µg/mL	Fe 100.0 ± 0.5 µg/mL
Manganese (Mn)	100.0 ± 0.5 µg/mL	Mn 100.0 ± 0.5 µg/mL
Nickel (Ni)	100.0 ± 0.5 µg/mL	Ni 100.0 ± 0.5 µg/mL
Vanadium (V)	100.0 ± 0.5 µg/mL	V 100.0 ± 0.5 µg/mL
Zinc (Zn)	100.0 ± 0.5 µg/mL	Zn 100.0 ± 0.5 µg/mL

FRM/82/11/4/18263 Rev No. 2 Date 24/07/2023

Prepared by Dnyaneshwar

In compliance with the standard specifications of SRL

Plant Site 2 : H-4, MIDC, Taloja – 410 208, Maharashtra, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.