



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product **10935 - Soyabean Casein Digest Medium (SCDM)
(Tryptone Soya Broth, TSB), Harmonized**

Batch No **3415522**
Analysis Date 09-May-2024
Date of Manufacture May 2024
Expiry Date/Re-Test Date May 2027

Test Parameters	Standards	Actual Results
Appearance (Colour)	Light biege	Complies
Solubility	30.00 gm/liter	Complies
Solubility before autoclaving (Clarity)	Clear	Complies
pH (25°C)	7.3 ± 0.2	7.14 Complies
Prepared Medium Appearance after autoclaving (Clarity)	Clear	Complies
Prepared Medium Appearance after autoclaving (Colour)	Light amber	Complies
Cultural Response	Inoculate and incubate at specified temperatures for 18-24 hours or upto 7 days incase of fungi*.	
Organism	Staphylococcus aureus ATCC 25923	
Inoculum (cfu) 10-100	Growth : Good, Incubation temperature : 30 - 35°C	Complies
Organism	Bacteroides vulgatus ATCC 8482	
Inoculum (cfu) 10-100	Growth : Good Incubation temperature : 30 - 35°C	Complies
Organism	Staphylococcus aureus ATCC 6538P	
Inoculum (cfu) 10-100	Growth : Good Incubation temperature : 30 - 35°C	Complies
Organism	Staphylococcus epidermidis ATCC 12228	
Inoculum (cfu) 10-100	Growth : Good, Incubation temperature : 30 - 35°C	Complies
Organism	Candida albicans * ATCC 10231	
Inoculum (cfu) 10-100	Growth : Good Incubation temperature : 20 - 25°C	Complies
Organism	Streptococcus pyogenes ATCC 19615	
Inoculum (cfu) 10-100	Growth : Good, Incubation temperature : 30 - 35°C	Complies

FRM/82/11/3/10935 Rev No. 1 Date 19/07/2022

Prepared by Shikha

In compliance with the standard specifications of SRL

Plant Site 1 : D-88/2, MIDC, Turbhe – 400 705, New Mumbai, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/2

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](https://www.iso.org/standard/68549.html)

This COA has been generated electronically and it is valid without signature.



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product

**10935 - Soyabean Casein Digest Medium (SCDM)
(Tryptone Soya Broth, TSB), Harmonized**

Batch No

3415522

Analysis Date

09-May-2024

Date of Manufacture

May 2024

Expiry Date/Re-Test Date

May 2027

Test Parameters

Standards

Actual Results

Organism

Pseudomonas aeruginosa
ATCC 27853

Inoculum (cfu) 10-100

Growth : Good
Incubation temperature : 30 - 35°C

Complies

Organism

Streptococcus pneumoniae
ATCC 6305

Inoculum (cfu) 10-100

Growth : Good
Incubation temperature : 30 - 35°C

Complies

Organism

Micrococcus luteus
ATCC 9341

Inoculum (cfu) 10-100

Growth : Good,
Incubation temperature : 30 - 35°C

Complies

Organism

Salmonella typhimurium
ATCC 14028

Inoculum (cfu) 10-100

Growth : Good,
Incubation temperature : 30 - 35°C

Complies

FRM/82/11/3/10935 Rev No. 1 Date 19/07/2022

Prepared by Shikha

In compliance with the standard specifications of SRL

Plant Site 1 : D-88/2, MIDC, Turbhe – 400 705, New Mumbai, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

2/2

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.