



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product	73530 - 20X DoubleQuick DNA Electrophoresis Buffer
Batch No	9475880
Analysis Date	08-Jan-2024
Date of Manufacture	January 2024
Expiry Date/Re-Test Date	January 2025

Test Parameters	Standards	Actual Results
Appearance (Form)	Liquid	Liquid
Appearance (Colour)	Colourless	Colourless
Electrophoresis test	Electrophoresis of DNA samples in 0.8% (w/v) agarose gel at 20 V should resolve individual clear bands	Conforms
Visualization of bands after staining	Clear definite bands seen	Conforms
Gel strength	Preparation of 0.8% (w/v) agarose gel in 1x buffer. Should polymerize in 5-10 minutes.	Conforms

FRM/82/11/5/73530 Rev No. 1 Date 31/07/2021

Prepared by KalaiSelvi

In compliance with the standard specifications of SRL

Plant Site 3 : No. 21, Thomas Nagar, Little Mount, Chennai - 600 015, India.

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

This COA has been generated electronically and it is valid without signature.

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)